

Reference Values for Immature Granulocytes in Healthy Blood Donors Generated on the Sysmex XE-2100 Automated Hematology Analyser

M BRUEGEL*¹, GM FIEDLER*¹, G MATTHES*², and J THIERY*¹

*¹ Institute of Laboratory Medicine, Clinical Chemistry and Molecular Diagnostics, University Hospital Leipzig, Germany.

*² Institute of Transfusion Medicine, University Hospital Leipzig, Germany.

An increase in immature granulocytes (IG) in peripheral blood is of great importance for the diagnosis of hematological malignancies. Furthermore, it may indicate systemic inflammation. IGs are usually quantified by visual microscopy; however, recent studies have evaluated automated IG count on the DIFF channel and IG detection on the IMI channel of the Sysmex XE-2100 automated hematology analyser.

The aim of the present study was to generate reference values for these parameters. Men and women showed comparable reference intervals for the IG counts; the 5th percentile was zero and the 95th percentile was $0.03 \times 10^9 / L$ for both men and women. For the IG count in percent (%IG) the 5th percentile was zero and the 95th percentile was 0.5% for men and 0.4% for women. The count of myelocytes in the IMI channel in the study population resulted in a 5th percentile of zero and a 95th percentile of $0.03 \times 10^9 / L$ for both men and women.

The clinical value of these automated measurements has to be elucidated in further clinical studies.

(Sysmex J Int 14 : 5-7, 2004)

Key Words

Immature Granulocytes, IG Count, Immature Information, IMI, Reference Intervals

Received 7 May, 2004; Accepted 14 May, 2004

INTRODUCTION

An increase in immature granulocytes (IG) in peripheral blood is of great importance for the diagnosis of hematological malignancies. Furthermore, it may indicate systemic inflammation. Thus, automated IG counting may function as a screening test for infectious diseases and myeloid disorders, resulting in earlier diagnosis and intervention.

Recent studies have evaluated the automated IG count on the DIFF channel¹⁾ and IG detection by the IMI channel²⁾, but reference values for these two parameters have not been established until now.

The aim of the present study was to define reference values for the IG count on the DIFF channel and IG detection on the IMI channel of the Sysmex XE-2100 automated hematology analyser.

MATERIALS AND METHODS

Blood specimen selection and processing

156 healthy blood donors (80 men and 76 women: mean age 38 years [range 19 - 69 years]) from the Institute of Transfusion Medicine, University Hospital Leipzig, Germany, were enrolled in this study. Blood specimens

were collected into 2.8 mL K₃EDTA Monovettes (Sarstedt, Germany) and processed within 4 hours on the Sysmex XE-2100. Instrument performance was monitored using quality control materials supplied by the manufacturer.

All blood donors were assessed according to German legal requirements³⁾. This allowed to detect influencing and interfering events which might affect the rate of IG production including diet, physical activity, smoking, alcohol consumption and consumption of drugs or pharmaceuticals including oral contraceptives. Volunteers were also questioned about history of allergy, recent surgery, recent transfusion, bacterial infection and any other co-existing disease. Donors with C-reactive protein (CRP) levels > 5 mg/L were excluded from the study.

XE-2100 IG count

The IG count includes mainly metamyelocytes and myelocytes and is obtained from the DIFF channel by fluorescence flow cytometry. The combination of side scatter and fluorescence intensity of the nucleated cells characterizes each cell detected. The standard XE-2100 offers the IG count as a research parameter, however, upgrading with the XE-IG Master software is necessary for reporting this parameter⁴⁾.

Immature Information (IMI) channel measurement

The IMI (Immature Information) channel detects immature myeloid cells, including bands, IGs and blasts. The reaction principle of this channel is based on differences in membrane composition between mature and immature cells that are exploited by channel specific reagents resulting in a scattergram plotting radio frequency (RF y-axis) against direct current (DC x-axis). This channel determines the total number of myeloid precursor cells².

RESULTS

Measurement of IG count and IMI count on Sysmex XE-2100

A total of 156 samples was analyzed on the XE-2100 in the CBC+ DIFF mode. Results of the IG count are shown in **Fig. 1**: 26 blood donors showed a zero count; 12 had a count of $0.01 \times 10^9/L$; 78 a count of $0.02 \times 10^9/L$; 22 of $0.03 \times 10^9/L$; 11 of $0.04 \times 10^9/L$; 5 of $0.05 \times 10^9/L$; and 2 of $0.06 \times 10^9/L$.

Men and women showed comparable values for IGs with the highest value of $0.03 \times 10^9/L$ for men and $0.06 \times 10^9/L$ for women (**Figs. 2 and 3**). No age dependency for the IG counts was detectable (**Fig. 4**). The 5th percentile was zero and the 95th percentile was $0.03 \times 10^9/L$ for both men and women. For the IG count in percent (%IG) the 5th percentile was zero and the 95th percentile was 0.5% for men, 0.4% for women and 0.4% if the data for men and women were combined (data not shown).

The count of immature myelocytes in the IMI channel in this study population resulted in a 5th percentile of zero and a 95th percentile of $0.03 \times 10^9/L$ for both men and women (data not shown).




Fig. 1 Measurement of IG counts in healthy blood donors
The frequency of various concentrations is given in per cent.




Fig. 2 Measurement of IG counts in healthy men
The frequency of various concentrations is given in per cent.




Fig. 3 Measurement of IG counts in healthy women
The frequency of various concentrations is given in per cent.




Fig. 4 Age dependency of IG counts in healthy blood donors
(n = 156).

DISCUSSION

The aim of this study was to generate reference values for (1) the IG count and (2) IG detection by the IMI channel, analyzed by the Sysmex XE-2100 automated hematology analyser.

The automated analysis of IGs by IG count or IMI detection may be helpful in screening or monitoring leukaemoid reactions, such as severe and chronic infections, inflammation and tissue necrosis, neoplasia and myeloproliferative disorders. Some investigators recommend the IG values produced by the Sysmex XE-2100 as screening information about the infection status of the patient⁴⁾. Results from patients with IG flags were compared with the counts of myeloid precursors (promyelocytes, myelocytes, metamyelocytes) determined microscopically. It has been shown that IG counts on the Sysmex analyser were systematically lower than the manual counts but that both methods correlate acceptably⁵⁾.

A recent study attributes high sensitivity to the IMI channel for the detection of immature cells in pregnant women⁶⁾. Ansari-Lari, et al. state that > 3 % IGs, measured by Sysmex XE-2100, predict sepsis with > 90% specificity⁷⁾.

We analyzed IGs in 156 healthy blood donors by using IG count and IMI channel. We showed gender and age independent values.

Studies will be necessary to evaluate the application of automated IG measurements as screening assays for infectious diseases. However, reference values, generated in this study, should help to classify increased values of IGs.

CORRESPONDING AUTHOR:

Dr. Mathias Bruegel
Institute of Laboratory Medicine, Clinical Chemistry and
Molecular Diagnostics,
University Hospital Leipzig
Liebigstrasse 27
D-04103 Leipzig
Germany.
mathias.bruegel@medizin.uni-leipzig.de

References

- 1) Weiland TH, Kalkman H, Heihn H: Evaluation of the automated immature granulocyte count (IG) on Sysmex XE-2100 automated haematology analyser vs. visual microscopy (NCCLS H20-A). *Sysmex J Int*, 12: 63-70, 2002.
- 2) Herklotz R, Huber AR : Precision and accuracy of the leukocyte differential on the Sysmex XE-2100. *Sysmex J Int*, 11: 8-21, 2001.
- 3) Richtlinien zur Gewinnung von Blut- und Blutbestandteilen und zur Anwendung von Blutprodukten (Hämotherapie); Bundesgesundheitsbl.-Gesundheitsforsch.-Gesundheitsschutz; 43 : 555-589, 2000 .
- 4) Briggs C, et al.: New quantitative parameters on a recently introduced automated blood cell counter-the XE 2100. *Clin Lab Haem*, 22: 345-350, 2000.
- 5) Stamminger G, Auch D, Diem H, Sinha P: Performance of the XE-2100 leucocyte differential. *Clin Lab Haem*, 24: 271-280, 2002.
- 6) Fernandez-Suarez A, Pascual VT, Gimenez MTF, Hernandez JFS: Immature granulocyte detection by the SE-9000 haematology analyser during pregnancy. *Clin Lab Haem*, 25: 347-351, 2003.
- 7) Ansari-Lari MA, Kickler TS, Borowitz MJ: Immature granulocyte measurement using the Sysmex XE-2100. Relationship to infection and sepsis. *Am J Clin Pathol*, 120: 795-799, 2003.